

Available online at www.sciencedirect.com

Resuscitation Plus

RESUSCITATION COUNCIL

journal homepage: www.journals.elsevier.com/resuscitation-plus

Simulation and education

The efficacy and usability of suction-based airway clearance devices for foreign body airway obstruction: a manikin randomised crossover trial



Emma Patterson a,1, Ho Tsun Tang a,1, Chen Ji a, Gavin D. Perkins a,b, Keith Couper a,b,

Abstract

Background: Newly-developed suction-based airway clearance devices potentially provide a novel way to improve outcome in patients with foreign body airway obstruction. We conducted a randomised controlled crossover manikin trial to compare the efficacy and usability of two of these devices with abdominal thrusts.

Methods: We randomised participants from a UK medical school to one of six groups which determined the order in which participants attempted the three techniques (abdominal thrusts; LieVac, Nesconset, New York, USA; Dechoker, Concord North Carolina, USA). Randomisation was performed using an online randomisation system. Following brief training, participants sought to remove a foreign body alway obstruction from a manikin using the allocated technique. The primary outcome was successful removal of the foreign body. Usability was assessed in a questionnaire following the three simulations.

Results: We randomised and analysed data from 90 participants (58% male; 86% aged 18–29 years). Compared with abdominal thrusts, successful foreign body airway obstruction removal was achieved more frequently in manikins in the LifeVac group (odds ratio 47.32, 95% CI 6.75–389.40) but not in the Dechoker group (odds ratio 1.22, 95% CI 0.60–2.47). The usability of LifeVac and abdominal thrusts were generally evaluated more positively than the Dechoker.

Conclusion: In this manikin study, we found that, compared with abdominal thrusts, the success rate for foreign body airway obstruction removal was higher in the LifeVac group but not in the Dechoker group.

Keywords: Airway obstruction, Choking, Basic life support, Anti-choking device, Randomised controlled trial, Simulation

Introduction

Foreign body airway obstruction (FBAO) is an important cause of mortality and morbidity, particularly in the very young and old. ^{1,-2} Each year, FBAO is responsible for almost 2,000 ambulance calls in London and approximately 250 UK deaths. ^{1,3}

Current treatment for FBAO is based on a step-wise approach, that incorporates techniques including coughing, back blows, abdominal

thrusts, and chest thrusts/compressions.* Abdominal thrusts are reserved for severe cases of FBAO that are not relieved by back blows, due to associated risk of thoracic, vascular and gastrooesophageal injury.* Evdence supporting specific interventions is limited, such that current treatment recommendations are based predominantly on case series and expert opinion.*.6

The risks associated with current treatments for FBAO have driven interest in alternative strategies for FBAO removal. In recent years, new suction-based airway clearance devices have been developed in

http://dx.doi.org/10.1016/j.resplu.2020.100067

Received 11 December 2020; Accepted 13 December 2020

Available online xxx

^a Warwick Medical School, University of Warwick, Coventry, UK

^b Critical Care Unit, University Hospitals Birmingham NHS Foundation Trust, Birmingham, UK

^{*} Corresponding author.

E-mail address: k.couper@warwick.ac.uk (K. Couper).

Joint first author.

which manual suction is applied via a face mask to relieve FBAO. A recent systematic review of these devices of sentified published data for only one device. Available studies for this device were limited to manikin studies, cadaver studies, and clinical case series. Based on the limited data published to date, the International Liaison Committee on Resuscitation has decided that it would be premature to make a recommendation for or against the use of devices, and highlighted the urgent need for futher research.

To date, no study has compared these devices with standard care. The efficacy and usability of new devices, in comparison with standard care, are important factors in determining whether a medical device should be adopted in practice. In view of the current absence of evidence in relation to this important issue, we identified the specific need for research in this area.

Methods

We conducted an open-label, randomised controlled crossover manikin trial to compare the efficacy and usability of two suctionbased airway clearance devices (LifeVac, Nesconset, New York, USA; Dechoker, Concord, North Carolina, USA) with the abdominal thrust.

The LifeVac comprises a facemask attached to compressible bellows. To use the device, the mask is held over the choking patent's mouth and nose, and then the handle of the bellows is pressed downwards and sharply pulled upwards. The Dechoker comprises a facemask attached to an oropharyngeal tube attached to a large cylinder with a plunger. To generate negative pressure, the plunger is pulled backwards sharply. Both devices are promoted as being straightforward to use. 10.11

The trial protocol was finalised before the start of the study. The study was reviewed and approved by the University of Warwick Biomedical & Scientific Research Ethics Committee (reference 108/ 18-19). Written informed consent was obtained from all participants. No changes were made to the trial protocol following commencement.

Setting and participants

The study was conducted in the Medical School at the University of Warwick. We included university staff and students that could communicate in English and who provided written informed consent to participate. We excluded individuals who had a physical disability that precluded use of the devices.

Randomisation

Following confirmation of eligibility and provision of written informed consent we randomised participants in an equal ratio to one of six groups that determined the order in which they completed the three interventions. Details of the groups and corresponding order are included in figure one and the electronic Supplement (Table S1). The randomisation sequence was developed using an online system using a fixed block size of six by a researcher that was not involved in participant recruitment. Terror randomisation, we used an online randomisation system to maintain allocation concealment. Following randomisation, participants were informed only of the intervention that they would be requested to complete next in the sequence.

Interventions and study process

The researcher showed the participant a short information video on how to deliver the first intervention. For the LifeVac and Dechoker, we extracted key information from manufacturer training videos freely available on the internet. ^{10,11} For abdominal thrusts, we extracted information from a video on foreign body airway obstruction developed by a UK first aid charity. ¹⁴ Participants were not given the opportunity to handle the device or practice any technique prior to the simulated scenario.

For the scenario, participants were informed that a 25-year old male was eating steak at a restaurant when they suddenly began to cough and pointing to their throat. Back slaps had been attempted, but these were ineffective. For the patient, we used a manikin (Choking Charlie, Laerdal Medical AS, Stavanger, Norway) with a simulated food bolus sited in the manikin's throat, as per manufacturer instructions. The participant was then to perform the allocated intervention. To ensure consistency across interventions, participants were permitted only to use the allocated intervention. Participants were over our to four-minutes to remove the obstruction.

After the first scenario, we adopted the same procedure for subsequent interventions. There was no break between attempting interventions. Following scenario three, participants completed a questionnaire on device usability. It was not possible to blind either the research participant or outcome assessor for teatment allocation.

Outcomes

The primary study outcome was successful removal of the foreign body airway obstruction within four-minutes. This was defined as the removal of the simulated food bolus from the manikin's mouth. The four-minute period was timed by a single researcher with a stonwatch.

The secondary efficacy outcome was time to FBAO removal. A single researcher present during the scenario measured the time in seconds from the start of the scenario to the point that the FBAO exited the manikin's mouth using a stopwatch. Secondary usability outcomes were captured in a survey completed at the end of the three scenarios. For each device, participants were asked to rank five statements on a scale of 1 strongly disagrel to 10 (strongly agree). These statements were: Lunderstood how to use the device, the device was easy to learn; the device was easy to use; I felt confident using this device; and I would feel confident using this device in a real-tile emergency.

Sample size

We selected a sample size of 90 participants. In the absence of any preliminary data to provide insights in to expected effect size, our sample size was chosen based on the time frame available for data collection and the size of the pool of potential participants.

Statistical methods

We describe categorical data as number and frequency. We describe all continuous data as median and interquantile range to reflect the type of data collected. For our primary outcome (successful removal), we list assessed for a group, period or carryover effect, using a mixed-effects binary logistic regression model. In the absence of such effects, we used the same model framework to estimate the effect in

removing the foreign body airway obstruction for both LifeVac and Dechoker, compared with abdominal thrusts. Participants were included as a random-effect in the model. The analysis was not adjusted for any covariates.

For time to removal, we visualised data using a Kaplan-Meier sunvival curve. As indicated by the crossed curves, violation of the proportional hazards assumption precluded use of a cox proportional hazards assumption precluded use of a cox proportional hazards model or ordinal regression. Weighted log-rank tests were not used as the crosses occurred at different time points. The proportional odds assumption was assessed by the test of parallel lines. As such, we categorised time to removal in to five groups based on time to removal (group 1: 0 - 59 seconds, group 2: 60 - 119 seconds, group 3: 120 - 179 seconds, group 4: 180 - 239 seconds, and group 5: not successfully removed). We then adopted the same modelling strategy described for our primary outcome to compare groupings (group one val other groups; groups one) we dill other groups; groups one/two val other groups, etc).

For usability outcomes, we compared across all three groups using Friedman's test. In the event that the overall test was statistically significant (p < 0.05), we compared differences between pairs of groups (LifeYac v Abdominal thrusts; LifeYac v Dechoker; Dechoker v Abdominal thrusts) using the Wilcoxon signed-rank test.

The analyses were conducted on a per-protocol basis. We present model results as odds ratio and 95% confidence interval (CI) and reported p values for the non-parametric test results. All primary statistical tests were two-sided with a pre-specified significance level of 0.5. Pairwise comparisons of the usability outcomes were two-sided with a Bonferroni correction applied to account for multiple testing, such that pairwise level of significance was 0.107 (0.05 divided by three). We underlook analyses using SPSS (version 26.0, IBM Corp. Armonk, New York) and STATA (version 16.0, StataCorp. College Station, Texas).


Results

In October 2019, 93 individuals were screened for study participation, of which 92 participants were eligible, provided written informed consent and were randomised (Fig. 1). In two cases, participants did not complete all three tests correctly, such that they were not included in the analysis. Data from 90 individuals were available for analysis.

Most participants were male (n = \$2, 58%), aged 18 – 29 (n = 77, 68%), and a medical student (n = 86, 96%) (Table 1). Most participants had previously attended a first aid course (n = 85, 94%). Few participants had previously seen a LifeVac or Dechoker device. Participant characteristics were similar across the study groups (Supplementary appendix Table S2).

For the primary outcome, the FBAO was successfully removed in 99% cases with LIGH2ac, 74% cases with Dechoker, and 71% cases with abdominal thrusts (Table 2). The odds of successful removal was significantly higher in the LIGH2ac group than abdominal thrusts (odds ratio 47.32, 99% CI 575–389.40), but was not significantly higher in the Dechoker group compared with abdominal thrusts (odds ratio 1.22, 95% CI 0.60–2.47).

For time to removal, Fig. 2 shows the timing of success across groups. The crossed curves indicate the violation of proportional hazards assumption. Removal in less than one-minute occurred in 82% cases using LifeVac, 44% cases using Dechoker and 67% using abdominal thrusts. After the first minute, the FBAO was successfully removed in 17% cases using LifeVac, 30% cases using Dechoker, and 4% cases using abdominal thrusts. Across group comparisons, Lifevac was consistently superior to abdominal thrusts. For Dechoker, comparison of group one (removal in less than one minute) with subsecuent time periods showed Dechoker to be less efficacious than



Notation under group name indicates order of interventions: AT-Abdominal thrusts; LV-LifeVac; DC-Dechoker Reasons for post-randomisation exclusions: 1- one individuals did not follow standardised procedure for foreign body airway obstruction removal; ±1- device broke during scenario.

Fig. 1 - CONSORT participant flow diagram.

Table 1 - Participant characteristics.

a One participant declined to answer.

	All (n = 90)
Age (years)-n(%) ^a	Turing and the
18-29	77 (85.6%)
30-39	8 (8.9%)
40-49	2 (2.2%)
50-59	2 (2.2%)
Sex- male-n (%) [®]	52 (58.4%)
Role- n (%)	
Student-medical	86 (95.6%)
Student-other	0 (0%)
Staff	4 (4.4%)
Attended first aid course- Yes-n (%)	85 (94.4%)
Real-life experience of FBAO management-n (%)	
None	72 (80.0%)
Back slaps	15 (16.7%)
Back slaps/abdominal thrusts	3 (3.3%)
Previously seen Life-Vac-n (%)	6 (6.7%)
Previously seen Dechoker-n (%)	3 (3.3%)

abdominal thrusts (odds ratio 0.38, 95% CI 0.20 to 0.72). This effect was not observed in subsequent time point comparisons.

Participants reported that they understood how to use all three techniques (Table 3). For all other usability outcomes, we observed statistically significant differences across the three groups. The LifeVac consistently outperformed the Dechoker device, whilst comparisons between the other two groups (LifeVac v Abdominal thrusts: Dechoker v Abdominal thrusts) were mixed. Reported confidence using techniques in real-life was highest in the abdominal thrust group, although between group comparisons showed abdominal thrusts were not superior to the LifeVac.

Discussion

In this manikin randomised crossover trial of 90 participants, we identified that use of LifeVac resulted in both quicker FBAO removal and greater overall success. Dechoker was not superior to abdominal thrusts. Success rates in the LifeVac group were reflected across usability outcomes.




Fig. 2 - Time to removal of foreign body for study interventions.

The successful removal of the FBAO without harm to the patient is the primary aim of all FBAO treatments. Following their first description in 1974 and despite early controversy, abdominal thrusts have become a core component of FBAO guidelines. 4,15,16 However, abdominal thrust success rates are challenging to determine as data are limited to case series. In our study, a population of predominantly medical students that had previously undertaken a first aid course achieved a success rate of 71%. The most robust clinical report of abdominal thrusts effectiveness reported a FBAO removal success rate of 79%, although this is likely an over-estimate due to selection bias and recall bias. 15 In contrast to suction-based airway clearance devices, a key advantage of abdominal thrusts is that they require no additional equipment to perform. Modifications have been described for use in patients that are unable to stand.17

For the two devices (LifeVac and Dechoker), published data on success rates are very limited.7 A systematic review identified no published peer-reviewed studies of the Dechoker device.7 In a manikin study of LifeVac, participants achieved a 94% success rate with one attempt and a 100% success rate with three attempts. 18 A cadaver study of LifeVac reported a 98% success rate on the first attempt, and a 100% success rate with two attempts. 19 The overall success rate for the LifeVac of 99% in our study is broadly consistent with these previous studies.

A key issue with these devices is that their use may distract the rescuer from other techniques, such as back slaps, abdominal thrusts and chest thrusts. The successful removal of an FBAO using devices

Table 2 - Study outcomes.

				Between group comparisons (odds ratio (95% confidence interval))			
	LifeVac	Dechoker	Abdominal thrust	LifeVac v abdominal thrusts	Dechoker v abdominal thrusts		
FBAO removal success-n (%)	89 (98.9%)	67 (74.4%)	64 (71.1%)	47.32 (5.75-389.40)	1.22 (0.60-2.47)		
Time to removal- n (%)							
Group 1: 0-59 seconds	74 (82.2%)	40 (44.4%)	60 (66.7%)	2.39a (1.17-4.88)	0.38" (0.20 - 0.72)		
Group 2: 60-119 seconds	13 (14.4%)	14 (15.6%)	2 (2.2%)	13.53 ^b (3.83-47.86)	0.67 ^b (0.36-1.25)		
Group 3: 120-179 seconds	1 (1.1%)	6 (6.7%)	1 (1.1%)	24.95° (5.17-120.50)	0.83° (0.42-1.65)		
Group 4: 180-239 seconds	1 (1.1%)	7 (7.8%)	1 (1.1%)	47.32d (5.75-389.40)	1.22 ^d (0.60-2.47)		
Unsuccessful (Group five)	1 (1.1%)	23 (25.6%)	26 (28.9%)				
* Comparison of group 1 v groups 2-5.							

^b Comparison of groups 1-2 v groups 3-5.

^c Comparison of groups 1-3 v groups 4-5.

d Comparison of groups 1-4 v group 5.

Table 3 - usability outcomes.

	LifeVac median (IQR)	Dechoker median (IQR)	Abdominal thrust median (IQR)	p-value ^a	P-value for comparison between groups ^b		
					LifeVac v Dechoker	LifeVac v abdominal thrusts	Dechoker v abdominal thrusts
Understand how to use technique	9.0 (7.0-10.0)	9.0 (7.0-10.0)	9.0 (8.0-10.0)	0.115		-	-
Technique easy to lean	9.0 (8.0-10.0)	8.0 (6.0-9.0)	9.0 (7.0-10.0)	< 0.001	0.007	0.47	0.015
Technique easy to use	9.0 (6.0-10.0)	6.0 (4.0-8.3)	7.0 (5.0-9.0)	< 0.001	< 0.001	0.013	0.08
Confident using technique	8 (6.0-9.0)	6.0 (2.0-8.0)	7.5 (5.0-9.0)	< 0.001	< 0.001	0.50	< 0.001
Confidence using technique in real-life emergency	7.0 (5.5-9.0)	5.0 (1.0-8.0)	8.0 (5.0-9.0)	<0.001	<0.001	0.84	<0.001

IQR, interquartile range

a p-values based on 90 comparisons except confidence using technique in real-life emergency (89 comparisons).

b p-values based on 90 comparisons except confidence using technique in real-life emergency-LifeVac v Dechoker (89 comparisons); confidence using technique in real-life emergency-DeChoker v Abdominal thrusts (89 comparisons).

relies on the generation of sufficient negative pressure, which is dependent on achieving an effective facemask seal. Previous research highlights the challenge of achieving an adequate seal with a face mask, particularly when using a one-handed technique.^{20–22} Our study recruited in a medical school such that most participants were medical students and may have a creater awareness of the

Our study recruited in a medical school such that most participants were medical students and may have a greater awareness of the importance and technique for generating an adequate seal than the general public.

The key difference between the Dechoker and LifeVac is that the

DeChoker incorporates an oropharyngeal tube. Theoretically, the tube should focus the generated negative pressure to a specific location to facilitate FBAO removal. However, in our study, the LifeVac was superior to the Dechoker both in terms of overall success rates and time to removal. In the clinical setting, an important concern is that the insertion of the orophrangeal tube component of the Dechoker has parallels with a blind finger sweep, which are associated with harms such as soft tissue injury and the risk of inadvertent FBAO translocation making it more difficult to remove.^{23–24}

Our study has a number of important limitations. Firstly, mankins studies provide an important way to test the efficacy of FBAO interventions using standardised processes. However, general-isability to the clinical setting is limited as it is not possible to recreate the fidelity of a time-critical clinical event. Secondly, our simulated obstruction was a small hard spherical object. Performance of different techniques will likely any with obstructions of different consistencies and size. Thirdly, we recruited participants from a medical school which is reflected in the demographics of participants including the high proportion that had previously attended a first aid course. This may not be reflective of the general population. Fourthly, we were unable to blind either study participants or outcome assessors, which may have contributed to performance or detection bias.

Fifthy, the training for each intervention was relatively brief and did not allow participants the opportunity to practice. We used key components of manufacturer training information in our participant training videos. Based on this training, participants reported that they understood how to use study techniques. It is not known whether additional, more intense training may have influenced study results. Finally, we asked participants to continue using the same technique for the four-minute scenario. In contrast, clinical guidelines recommend alternating techniques if a specific technique does not quickly lead to successful FBAO removal.⁴

Conclusion

In this manikin study, we found evidence that individuals using the LifeVac were more successful in removing a simulated foreign body airway obstruction than individuals using abdominal thrusts. We did not find evidence of improved success by individuals using the Dechoker, compared with individuals using adominal thrusts. Further research in the clinical setting is needed to understand the potential role of suction-based airway clearance devices in the management of FBAO.

Funding

GDP is supported as an NIHR Senior Investigator and by the NIHR Applied Research Centre (ARC) West Midlands, UK. The views expressed are those of the author(s) and not necessarily those of the NIHR or the Department of Health and Social Care.

Conflict of interests

KC is an associate editor of Resuscitation Plus. The remaining authors have no conflicts of interest to declare.

CRediT authorship contribution statement

Emma Patterson: Methodology, Formal analysis, Investigation, Writing - original draft, Writing - review & editing, Ho Tsun Tang: Methodology, Formal analysis, Investigation, Writing - original draft, Writing- review & editing, Chen Ji: Formal analysis, Writing- review & editing, Supervision, Gavin D. Perfins: Conceptualization, Methodology, Formal analysis, Resources, Writing - review & editing, Supervision Meth Couper: Conceptualization, Methodology, Formal

analysis, Resources, Writing - original draft, Writing - review & editing, Visualization, Supervision.

Appendix A. Supplementary data

Supplementary material related to this article can be found, in the online version, at doi:https://doi.org/10.1016/j.resplu.2020.100067.

REFERENCES

- Pavitt MJ, Nevett J, Swanton LL, et al. London ambulance source data on choking incidence for the calendar year 2016: an observational study. BMJ Open Respir Res. 2017;4:e000215.
- Injury Facts. Preventable death and death rates per 100,000
 population in the home and community by cause and age group,
 United States, 2017. [Available from: https://injurylacts.nsc.org/homeand-community-home-and-community-overview/deaths-in-thehome-and-community-by-age-group-and-cause/, accessed 14th
 January 2020].
- Office for National Statistics. Choking related deaths registered in England and Wales, 2014 to 2017. [Available from https://www.ons. gov.uk/peoplepopulationandcommunity/birthsdeathsandmarriages/ deaths/adhocs/009342chokingrelateddeathsregisteredinengland
- andwales2014to2017, accessed 2nd November 2020].
 4. Perkins GD, Handley AJ, Koster RW, et al. European Resuscitation Council Guidelines for Resuscitation 2015: Section 2. Adult basic life support and automated external delibrillation. Resuscitation
- Sp5:81–99.
 Couper K, Abu Hassan A, Ohri V, et al. Removal of foreign body airway obstruction: A systematic review of interventions. Resuscitation. 2020;156:174–81.
- Olasveengen TM, Mancini ME, Perkins GD, et al. Adult basic life support: international consensus on cardiopulmonary resuscitation and emergency cardiovascular care science with treatment recommendations. Resuscitation. 2020;156:A35–79.
- Dunne CL, Peden AE, Queiroga AC, Gomez Gonzalez C, Valesco B, Szpilman D. A systematic review on the effectiveness of anti-choking suction devices and identification of research gaps. Resuscitation. 2020;153:219–26.
- LifeVac Europe Ltd. General information: LifeVac Europe Ltd; [Available from: https://www.lifevac.eu/lifevac-pamphlet.pdf, accessed 9th September 2020].

- Dechoker. How to Use Dechoker Anti-Choking Device (ACD)
 [Available from: https://www.dechoker.com/pages/dechoker-usage-
- instructions, accessed 9th September 2020].

 10. LifeVac. LifeVac Training 2018 [Available from: https://www.youtube.
- com/watch?v=aqizWSL2ndA, accessed 9th September 2020].
 Dechoker, How it Works: Dechoker, Anti-Choking Device 2018 [Available from: https://www.youtube.com/watch?v=MYVyoBrAlK4, accessed 9th September 2020].
- Sealed Envelope Ltd. Create a blocked randomisation list. [Available from: https://www.sealedenvelope.com/simple-randomiser/v1/lists, accessed 2nd November 2020].
 Sealed Envelope Ltd. Simple randomisation engine. [Available from:
- Sealed Envelope Ltd. Simple randomisation service. [Available from: https://www.sealedenvelope.com/simple-randomiser/v1/, accessed 2nd November 2020].
- St John Ambulance. What To Do When Someone Is Choking First Aid Training - St John Ambulance [Available from: https://www.youtube. com/watch?v=PA9hpOnvtCk accessed 9th September 2020].
- Redding JS. The choking controversy: critique of evidence on the Heimlich maneuver. Crit Care Med. 1979;7:475–9.
- Heimlich H. Pop goes the cafe coronary. Emerg Med. 1974;6:154–5.
 Blain H, Bonnafous M, Grovalet N, Jonquet O, David M. The table maneuver: a procedure used with success in four cases of unconscious choking older subjects. Am J Med 2010;123(1150):e7
- Lih-Brody L, Singer M, Brody E. 382 Lifevac: a novel device for the resuscitation of the adolescent choking victim. Ann Emergency Med. 2017;70:S149

 –50.
- Juliano M, Domingo R, Mooney MS, Trupiano A. Assessment of the LifeVac, an anti-choking device, on a human cadaver with complete airway obstruction. Am J Emergency Med 2016;34:1673

 –4.
- Otten D, Liao MM, Wolken R, et al. Comparison of bag-valve-mask hand-sealing techniques in a simulated model. Ann Emergency Med 2014;63:6—12 e3.
- Grastein NS, Carey MC, Braude DA, et al. Efficacy of facemask ventilation techniques in novice providers. J Clin Anesthesia. 2013;25:193—7.
- Hart D, Reardon R, Ward C, Miner J. Face mask ventilation: a comparison of three techniques. J Emergency Med 2013;44:1028
 33
- Abder-Rahman HA. Infants choking following blind finger sweep. J Pediatr 2009;85:273–5.
- Hartrey R, Bingham RM. Pharyngeal trauma as a result of blind finger sweeps in the choking child. J Accid Emergency Med 1995;12:52–4.
- Gjoni D, Mbamalu D, Banerjee A, James KK. An unusual complication of an attempt to open the airway in a choking child. Br J Hospital Med 2009;70:595.